

Dataset Expocode	33GG20191104
Primary Contact	Name: Sullivan, Kevin Organization: NOAA/AOML CIMAS Address: 4301 Rickenbacker Causeway, Miami, FI 33149 Phone: (305) 361-4382 Email: kevin.sullivan@noaa.gov
Investigator	Name: Wanninkhof, Rik Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory Address: 4301 Rickenbacker Causeway, Miami FI, 33149 Phone: 305-361-4379 Email: Rik.Wanninkhof@noaa.gov
Investigator	Name: Pierrot, Denis Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory Address: 4301 Rickenbacker Causeway, Miami FI, 33149 Phone: 305-361-4441 Email: Denis.Pierrot@noaa.gov
Dataset	Funding Info: NOAA Climate Program Office; NOAA Ocean Acidification Program Initial Submission (yyyymmdd): 20191119 Revised Submission (yyyymmdd): 20191119
Campaign/Cruise	Expocode: 33GG20191104 Campaign/Cruise Name: GU1907-transit Campaign/Cruise Info: AOML_SOOP_CO2 Platform Type: CO2 Instrument Type: Equilibrator-IR or CRDS or GC Survey Type: Research Cruise Vessel Name: R/V Gordon Gunter Vessel Owner: NOAA Vessel Code: 33GG
Coverage	Start Date (yyyymmdd): 20191104 End Date (yyyymmdd): 20191112 Westernmost Longitude: 83.2 W Easternmost Longitude: 71.4 W Northernmost Latitude: 41.4 N Southernmost Latitude: 24.2 N Port of Call: Newport, RI Port of Call: Tampa, FL
Variable	Name: xCO2_EQU_ppm Unit: ppm Description: Mole fraction of CO2 in the equilibrator headspace (dry) at equilibrator temperature (ppm)
Variable	Name: xCO2_ATM_ppm Unit: ppm Description: Mole fraction of CO2 measured in dry outside air (ppm)
Variable	Name: xCO2_ATM_interpolated_ppm Unit: ppm Description: Mole fraction of CO2 in outside air associated with each water analysis. These values are interpolated between the bracketing averaged good xCO2_ATM analyses (ppm)

Variable	Name: PRES_EQU_hPa Unit: hPa Description: Barometric pressure in the equilibrator headspace (hPa)
Variable	Name: PRES_ATM@SSP_hPa Unit: hPa Description: Barometric pressure measured outside, corrected to sea level (hPa)
Variable	Name: TEMP_EQU_C Unit: Degree C Description: Water temperature in equilibrator (°C)
Variable	Name: SST_C Unit: Degree C Description: Sea surface temperature (°C)
Variable	Name: SAL_permil Unit: ppt Description: Sea surface salinity on Practical Salinity Scale (o/oo)
Variable	Name: fCO2_SW@SST_uatm Unit: µatm Description: Fugacity of CO2 in sea water at SST and 100% humidity (µatm)
Variable	Name: fCO2_ATM_interpolated_uatm Unit: µatm Description: Fugacity of CO2 in air corresponding to the interpolated xCO2 at SST and 100% humidity (µatm)
Variable	Name: dfCO2_uatm Unit: µatm Description: Sea water fCO2 minus interpolated air fCO2 (µatm)
Variable	Name: WOCE_QC_FLAG Unit: None Description: Quality control flag for fCO2 values (2=good, 3=questionable)
Variable	Name: QC_SUBFLAG Unit: None Description: Quality control subflag for fCO2 values, provides explanation when QC flag=3
Sea Surface Temperature	Location: In engine room, about 2 m after the seachest, before the SW pumps. Manufacturer: Seabird, Inc. Model: SBE 38 Accuracy: 0.001 (°C if units not given) Precision: 0.0003 (°C if units not given) Calibration: Factory calibration Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.
Sea Surface Salinity	Location: In Chem lab, next to CO2 system Manufacturer: Seabird Model: SBE 45 Accuracy: ± 0.005 o/oo Precision: 0.0002 o/oo Calibration: Factory calibration Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

Atmospheric Pressure

Location: Next to the bridge, ~15 m above the sea surface water
Normalized to Sea Level: yes
Manufacturer: RMYoung
Model: 61201
Accuracy: ± 0.5 hPa (hPa if units not given)
Precision: 0.01 hPa (hPa if units not given)
Calibration: Factory calibration
Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

Atmospheric CO2

Measured/Frequency: Yes, 5 readings in a group every 3 hours
Intake Location: Bow mast, ~18 meters above sea surface
Drying Method: Gas stream passes through a thermoelectric condenser (~5 °C) and then through a Perma Pure (Nafion) dryer before reaching the analyzer (90% dry).
Atmospheric CO2 Accuracy: ± 0.5 μ atm in fCO2_ATM
Atmospheric CO2 Precision: ± 0.01 μ atm in fCO2_ATM

Aqueous CO2 Equilibrator Design

System Manufacturer:
Intake Depth: 5 meters
Intake Location: Bow
Equilibration Type: Spray head above dynamic pool, no thermal jacket
Equilibrator Volume (L): 0.95 L (0.4 L water, 0.55 L headspace)
Headspace Gas Flow Rate (ml/min): 70 - 150 ml/min
Equilibrator Water Flow Rate (L/min): 1.5 - 2.0 L/min
Equilibrator Vented: Yes
Equilibration Comments: Primary equilibrator is vented through a secondary equilibrator.
Drying Method: Gas stream passes through a thermoelectric condenser (~5 °C) and then through a Perma Pure (Nafion) dryer before reaching the analyzer (90% dry).

Aqueous CO2 Sensor Details

Measurement Method: IR
Method details: details of CO2 sensing (not required)
Manufacturer: LI-COR
Model: 6262
Measured CO2 Values: xco2(dry)
Measurement Frequency: Every 140 seconds, except during calibration
Aqueous CO2 Accuracy: ± 2 μ atm in fCO2_SW
Aqueous CO2 Precision: ± 0.01 μ atm in fCO2_SW
Sensor Calibrations:
Calibration of Calibration Gases: The analyzer is calibrated every 4 hours with field standards that in turn were calibrated with primary standards that are directly traceable to the WMO X2007 scale. The zero gas is ultra-high purity air.
Number Non-Zero Gas Standards: 3
Calibration Gases:

Std 1: LL100000, 0.00 ppm, owned by AOML, used every ~4.5 hours.
Std 2: JA02140, 234.21 ppm, owned by AOML, used every ~4.5 hours.
Std 3: JA02689, 406.90 ppm, owned by AOML, used every ~4.5 hours.
Std 4: JB03276, 471.65 ppm, owned by AOML, used every ~4.5 hours.
Comparison to Other CO2 Analyses:
Comments:
Method Reference:

Pierrot, D., C. Neil, K. Sullivan, R. Castle, R. Wanninkhof, H. Lueger, T. Johannessen, A. Olsen, R. A. Feely, and C. E. Cosca (2009), Recommendations for autonomous underway pCO₂ measuring systems and data reduction routines, Deep-Sea Res II, 56, 512-522.

Equilibrator

Location: Inserted into equilibrator ~5 cm below water level

Temperature Sensor

Manufacturer: Hart

Model: 1521

Accuracy: 0.025 (°C if units not given)

Precision: 0.001 (°C if units not given)

Calibration: Factory calibration

Comments: Resolution is taken as Precision.

Equilibrator

Location: Attached to equilibrator headspace.

Pressure Sensor

Manufacturer: Setra

Model: 270

Accuracy: 0.05 (hPa if units not given)

Precision: 0.015 (hPa if units not given)

Calibration: Factory calibration

Comments: Manufacturer's Resolution is taken as Precision.

**Additional
Information**

Suggested QC flag from Data Provider: NA

Additional Comments: The analytical system operated well during this cruise. The GPS system was worked on during this transit; missing position data was interpolated from surrounding good data. When more than one analysis in a row had poor GPS data, the analyses were flagged 3. Original Data Location: http://www.aoml.noaa.gov/ocd/ocdweb/gunter/gunter_introduction.html Full unprocessed data files from analytical instrument including flow information and ship's meteorological and TSG data at time of sampling can be obtained upon request.

Citation for this Dataset:

Other References for this Dataset: