

Dataset Expocode	33WA20191013
Primary Contact	Name: Sullivan, Kevin Organization: NOAA/AOML CIMAS Address: 4301 Rickenbacker Causeway, Miami, FL 33149 Phone: (305) 361-4382 Email: kevin.sullivan@noaa.gov
Investigator	Name: Rodriguez, Carmen Organization: RSMAS/University of Miami Address: 4600 Rickenbacker Causeway, Miami, FL 33149 Phone: 305-421-4708 Email: crodriguez@rsmas.miami.edu
Investigator	Name: Millero, Frank Organization: RSMAS/University of Miami Address: 4600 Rickenbacker Causeway, Miami FL, 33149 Phone: 305-421-4707 Email: FMillero@rsmas.miami.edu
Investigator	Name: Pierrot, Denis Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory Address: 4301 Rickenbacker Causeway, Miami FL, 33149 Phone: 305-361-4441 Email: Denis.Pierrot@noaa.gov
Investigator	Name: Wanninkhof, Rik Organization: NOAA/Atlantic Oceanographic & Meteorological Laboratory Address: 4301 Rickenbacker Causeway, Miami FL, 33149 Phone: 305-361-4379 Email: Rik.Wanninkhof@noaa.gov
Dataset	Funding Info: NOAA Climate Program Office; NOAA Ocean Acidification Program Initial Submission (yyyymmdd): 20200808 Revised Submission (yyyymmdd): 20200808
Campaign/Cruise	Expocode: 33WA20191013 Campaign/Cruise Name: WS19286 Campaign/Cruise Info: SOOP_CO2 Platform Type: CO2 Instrument Type: Equilibrator-IR or CRDS or GC Survey Type: Research Cruise Vessel Name: R/V F.G. Walton Smith Vessel Owner: University of Miami Vessel Code: 33WA
Coverage	Start Date (yyyymmdd): 20191013 End Date (yyyymmdd): 20191013 Westernmost Longitude: 80.2 W Easternmost Longitude: 79.8 W Northernmost Latitude: 25.8 N Southernmost Latitude: 25.4 N Port of Call: Miami, FL, USA
Variable	Name: xCO2_EQU_ppm Unit:

Description: Mole fraction of CO₂ in the equilibrator headspace (dry) at equilibrator temperature (ppm)

Variable

Name: xCO₂_ATM_ppm

Unit:

Description: Mole fraction of CO₂ measured in dry outside air (ppm)

Variable

Name: xCO₂_ATM_interpolated_ppm

Unit:

Description: Mole fraction of CO₂ in outside air associated with each water analysis. These values are interpolated between the bracketing averaged good xCO₂_ATM analyses (ppm)

Variable

Name: PRES_EQU_hPa

Unit:

Description: Barometric pressure in the equilibrator headspace (hPa)

Variable

Name: PRES_ATM@SSP_hPa

Unit:

Description: Barometric pressure measured outside, corrected to sea level (hPa)

Variable

Name: TEMP_EQU_C

Unit:

Description: Water temperature in equilibrator (°C)

Variable

Name: SST_C

Unit:

Description: Sea surface temperature (°C)

Variable

Name: SAL_permil

Unit:

Description: Sea surface salinity on Practical Salinity Scale (o/oo)

Variable

Name: fCO₂_SW@SST_uatm

Unit:

Description: Fugacity of CO₂ in sea water at SST and 100% humidity (µatm)

Variable

Name: fCO₂_ATM_interpolated_uatm

Unit:

Description: Fugacity of CO₂ in air corresponding to the interpolated xCO₂ at SST and 100% humidity (µatm)

Variable

Name: dfCO₂_uatm

Unit:

Description: Sea water fCO₂ minus interpolated air fCO₂ (µatm)

Variable

Name: WOCE_QC_FLAG

Unit:

Description: Quality control flag for fCO₂ values (2=good, 3=questionable)

Variable

Name: QC_SUBFLAG

Unit:

Description: Quality control subflag for fCO₂ values, provides explanation when QC flag=3

**Sea Surface
Temperature**

Location: After sea water pump in the forward, port hull

Manufacturer: Seabird, Inc.

Model: SBE 38

Accuracy: 0.001 (°C if units not given)

Precision: 0.0003 (°C if units not given)

Calibration: Factory calibration
Comments: Manufacturer's Resolution is taken as Precision; Maintained by ship.

Sea Surface Salinity **Location:** Near the sea water pump in the forward, port hull.
Manufacturer: Seabird
Model: SBE 45
Accuracy: ± 0.005 o/oo
Precision: 0.0002 o/oo
Calibration: Factory calibration
Comments: Manufacturer's Resolution is taken as Precision

Atmospheric Pressure **Location:** On mast above bridge at ~13 m above sea surface.
Normalized to Sea Level: yes
Manufacturer: R.M. Young
Model: 61302
Accuracy: ± 0.3 hPa (hPa if units not given)
Precision: 0.1 hPa (hPa if units not given)
Calibration: Factory calibration
Comments: Manufacturer's Resolution is taken as Precision.

Atmospheric CO2 **Measured/Frequency:** Yes, 5 readings in a group every 4.5 hours
Intake Location: On mast above the bridge at ~13 meters above the sea surface
Drying Method: Gas stream passes through a thermoelectric condenser (~5 °C) and then through a Perma Pure (Nafion) dryer before reaching the analyzer (90% dry).
Atmospheric CO2 Accuracy: ± 0.5 μ atm in fCO₂_ATM
Atmospheric CO2 Precision: ± 0.01 μ atm in fCO₂_ATM

Aqueous CO2 Equilibrator Design **System Manufacturer:**
Intake Depth: 1.5 meters
Intake Location: Bow
Equilibration Type: Spray head above dynamic pool, with thermal jacket
Equilibrator Volume (L): 0.95 L (0.4 L water, 0.55 L headspace)
Headspace Gas Flow Rate (ml/min): 70 - 150 ml/min
Equilibrator Water Flow Rate (L/min): 1.5 - 2.0 L/min
Equilibrator Vented: Yes
Equilibration Comments: Primary equilibrator is vented through a secondary equilibrator.
Drying Method: Gas stream passes through a thermoelectric condenser (~5 °C) and then through a Perma Pure (Nafion) dryer before reaching the analyzer (90% dry).

Aqueous CO2 Sensor Details **Measurement Method:** IR
Method details: details of CO₂ sensing (not required)
Manufacturer: LI-COR
Model: 6262
Measured CO2 Values: xco₂(dry)
Measurement Frequency: Every 140 seconds, except during calibration
Aqueous CO2 Accuracy: ± 2 μ atm in fCO₂_SW
Aqueous CO2 Precision: ± 0.01 μ atm in fCO₂_SW
Sensor Calibrations:
Calibration of Calibration Gases: The analyzer is calibrated every ~4.5 hours using field standards that were calibrated with primary standards at AOML that are directly traceable to the WMO scale. Ultra-High Purity air (0.0 ppm CO₂) and the high standard are used to zero and span the LI-COR analyzer.

Number Non-Zero Gas Standards: 4**Calibration Gases:**

Std 1: FF4297, 213.54 ppm, owned by RSMAS, used every ~4.5 hours.

Std 2: FF42246, 382.17 ppm, owned by RSMAS, used every ~4.5 hours.

Std 3: FF36858, 607.03 ppm, owned by RSMAS, used every ~4.5 hours.

Std 4: FF3582, 1530.42 ppm, owned by RSMAS, used every ~4.5 hours.

Std 5: LL100000, 0.00 ppm, owned by AOML, used every ~NaN hours.

Comparison to Other CO₂ Analyses:

Comments: Instrument is located in an air-conditioned laboratory.

Method Reference:

Pierrot, D., C. Neil, K. Sullivan, R. Castle, R. Wanninkhof, H. Lueger, T. Johannessen, A. Olsen, R. A. Feely, and C. E. Cosca (2009), Recommendations for autonomous underway pCO₂ measuring systems and data reduction routines, Deep-Sea Res II, 56, 512-522.

**Equilibrator
Temperature Sensor**

Location: Inserted into equilibrator ~5 cm below water level

Manufacturer: Hart

Model: 1523

Accuracy: 0.015 (°C if units not given)

Precision: 0.001 (°C if units not given)

Calibration: Factory calibration

Comments: Resolution is taken as Precision.

**Equilibrator
Pressure Sensor**

Location: Attached to equilibrator headspace. Differential pressure reading from Setra 239 attached to the equilibrator headspace is added to the pressure reading from the LICOR, which is measured by an external Setra 270 connected to the exit of the analyzer.

Manufacturer: Setra

Model: 270

Accuracy: 0.15 (hPa if units not given)

Precision: 0.015 (hPa if units not given)

Calibration: Factory calibration

Comments: Manufacturer's Resolution is taken as Precision.

**Additional
Information**

Suggested QC flag from Data Provider: NA

Additional Comments: The analytical system operated well during the cruise. Original Data Location: http://www.aoml.noaa.gov/ocd/ocdweb/wsmith/wsmith_introduction.html Full unprocessed data files from analytical instrument including flow information plus meteorological and TSG data at time of sampling can be obtained upon request.

Citation for this Dataset:

Other References for this Dataset: